

# **Clinical Validation report of Novel Coronavirus (SARS-CoV-2) Antigen Rapid Test Cassette (Swab)**



**Product name:** Novel Coronavirus (SARS-CoV-2) Antigen

Rapid Test Cassette (Swab)

**Package Specification:** 25 tests/kit

**Manufacturer:** Hangzhou Realy Tech Co., Ltd

## **I. Clinical validation time**

This clinical evaluation was conducted from July 2020 to Aug 14th,2020.

## **II. Background information for clinical evaluation**

Since December 2019, world has successively discovered multiple cases of patients with new-type coronavirus pneumonia. With the spread of the epidemic, China and abroad have also been found. As an acute respiratory infectious disease, the disease has been included in the Class B infectious diseases stipulated in the Law of the People's Republic of China on the Prevention and Control of Infectious Diseases, and is managed as a Class A infectious disease. Based on the current epidemiological investigation, the incubation period is 1-14 days, mostly 3-7 days.

The main manifestations are fever, dry cough, and fatigue. A few patients have symptoms such as nasal congestion, runny nose, sore throat, myalgia and diarrhea. Severe patients usually have dyspnea and / or hypoxemia one week after the onset of symptoms, and severe patients can quickly progress to acute respiratory distress syndrome, septic shock, difficult to correct metabolic acidosis, coagulation dysfunction and multiple organ Functional failure, etc. It is worth noting that in the course of severe and critically ill patients, there may be moderate to low fever, even without obvious fever.

Mild patients showed only low fever, mild fatigue, and no pneumonia. Judging from the current cases, most patients have a good prognosis, and a few patients are critically ill. The elderly and those with chronic underlying disease have a better prognosis. Symptoms in children are relatively mild.

The Novel Coronavirus (SARS-Cov-2) Antigen Rapid Test Cassette (Swab) developed by our company can help diagnose whether patients are infected with the Novel Coronavirus. It has further enriched the detection methods of Novel Coronavirus, expanded the supply of detection reagents, and fully served the needs of epidemic prevention and control.

## **III. Test purposes**

The Novel Coronavirus (SARS-Cov-2) Antigen Rapid Test Cassette (Swab) produced by Hangzhou Realy Technology Co., Ltd. was used to verify the feasibility of clinical evaluation and the reliability of test results for Chinese subjects.

The purpose of research of the clinical test is to calculate the consistency percentage of negative/positive and the total consistency percentage and Kappa coefficient by statistically analyzing test results through comparative experimental research.

## **IV. Test design**

### **1. Test plan selection and reasons**

In vitro diagnostic reagents for testing and reference reagents were used to conduct comparative research tests on clinically suspected Novel Coronavirus Nasopharyngeal swab samples, and it was proved that the in vitro diagnostic reagents used in the test can achieve the expected assistance in infection of the Novel Coronavirus.

### **2. Sample volume required**

File No. MF-K511416D-0010

Version: 1.5

Effective date:2020-11-25

The total number of clinical trials of this product is not less than 100 cases. The samples is classified into the positive group and the negative group as per the test results of the reference product. Meanwhile, the samples shall be tested via the qualitative test strip tested and by reference product from the same patient and then the test results of the product tested and the reference product shall be compared, with statistical analysis being made.

### **3. Sample inclusion/exclusion certification.**

The positive group and negative group in this experiment are applicable to the following inclusion/exclusion criteria

#### **Positive group inclusion:**

PCR Test is positive;

CT test results and symptoms are clinically positive;

#### **Negative inclusion:**

PCR test is negative;

CT test results and symptoms are clinically negative;

#### **Sample collection, processing**

It is applicable to the diagnosis of the Novel coroinavirus from the samples of Nasopharyngeal swab. Use freshly collected samples for optimal test performance. Inadequate sample collection or improper sample handling may yield a false-negative result.

**Sample collection procedure:** Completely insert the sterilized swab supplied in this kit into the nasal basin, and swab several times to collect the epidermal cells of the mucus.

It is recommended to collect sample from Nasopharyngeal for more accurate results.

#### **Specimen preparation:**

- 1) Take out 1 bottle of Sample Extraction Buffer, remove the bottle cap, add all the extraction buffer into the extraction tube supplied in this kit, and put it on the tube stand.
- 2) Insert the swab into the extraction tube which contains Sample Extraction Buffer. Rotate the swab inside the tube using a circular motion to roll the side of the extraction tube so that liquid is expressed and reabsorbed from the swab, remove the swab. The extracted solution will be used as test sample.

### **4. In vitro diagnostic reagents and reference products for testing**

#### **5.1 Test in vitro diagnostic reagents**

**Name:** The Novel Coronavirus (SARS-CoV-2) Antigen Rapid Test Cassette (Swab)

**Specification:** 25 tests/kit

**REF:** K511416D

**LOT:** 202007046

**Expiry:** June,2022(Tentative)

**Storage Conditions:** Store in a dry place at 2-30°C, protected from light. After opening the inner package, the test card will become invalid due to moisture absorption. Please use it within 1 hour.

**Source:** Hangzhou Realy Tech Co.,Ltd

#### **5.2 Reference products**

**Name:** Novel Coronavirus(2019-nCoV) Nucleic Acid Diagnostic Kit (PCR-Fluorescence Probing)

**Manufacturer:** Sansure Biotech Inc.

**Limit of detection:** The limit of detection of this kit is 200 copies/mL

File No. MF-K511416D-0010

Version: 1.5

Effective date: 2020-11-25

**Storage Conditions:** Store in a dry place at 2-8°C, protected from light.

## V. Experiment method

1. Get the Swab specimens from patients in positive and negative groups.
2. Pre-process the swab samples according to the instructions of The Novel Coronavirus (SARS-CoV-2) Antigen Rapid Test Cassette (Swab), and label the samples randomly.
  - 2.1 Add 10 drops (about 0.3 ml) of the sample extraction buffer into the extraction tube.
  - 2.2 Place the swab specimen in the SARS-CoV-2 antigen Buffer. Rotate the swab for approximately 10 seconds while pressing the head against the inside of the tube to release the antigen in the swab.
  - 2.3 Remove the swab while squeezing the swab head against the inside of Buffer as you remove it to expel as much liquid as possible from the swab. Discard the swab in accordance with your biohazard waste disposal protocol.
  - 2.4 Screw on and tighten the cap onto the specimen collection tube, then shake the specimen collection tube vigorously to mix the specimen and the Buffer. Place the test device on a clean and level surface.
3. The operation steps of the in vitro diagnostic reagents for the test are as follows. For details, please refer to the product instruction manual:
  - 3.1 remove the test sample and required reagents from the storage conditions and equilibrate to room temperature (15-30°C).
  - 3.2 When preparing for testing, open the aluminum foil bag from the tear. Remove the test card and lay it flat on a horizontal table.
  - 3.3 Label the sample number on the test card.
  - 3.4 Add 3 drops of the solution (approx. 80ul) to the sample well and then start the timer.
  - 3.5 Time counting and interpret the results within 10 minutes.

Note: The detection steps need to be completed under protection against infection.

## VI. Statistical methods of statistical analysis of clinical research data

### A Methods evaluating clinical performance

Whether various indexes can reach the standards of clinical evaluation shall be judged by calculating the consistency percentage of negative/positive and the total consistency percentage in the test results of the product tested and the reference product, to validate the accuracy and applicability of the product in clinical applications. The product tested shall be subject to tests through the sample of different types, with statistics on the results. Meanwhile, different types of sample of the subjects shall be subject to determination by the product tested synchronously, and then the determination results of both shall be compared. The test results recorded shall be subject to statistical analysis upon completion of determination of all clinical samples, to calculate the consistency percentage of negative/positive and the total consistency percentage. Afterwards, equivalence of both shall be evaluated as per these statistical indexes

File No. MF-K511416D-0010

Version: 1.5

Effective date: 2020-11-25

## B Statistical method

The products launched on the market shall be subject to comparative study and evaluation.Kappa inspection: each sample shall be tested with the product tested and the reference product respectively, and then the consistency in statistical results of these two inspection methods shall be compared through Kappa inspection.

The data shall be subject to Kappa inspection and analysis and the Kappa coefficient shall be calculated.Favorable consistency can be proven if Kappa is  $>0.8$ . The consistency in test results of the product tested and the reference product is evaluated as per the evaluation standards.

## VII Standards of clinical evaluation

The coincidence rate shall be calculated by comparing with the reference product whose marketing is approved. The product performance shall meet the following requirements.

1)Coincidence rate of negative: the sample whose test results are negative for both the product tested and the reference product and the proportion in the sample whose test results are negative for the reference product shall be more than 95%.

2)Coincidence rate of positive: the sample whose test results are positive for both the product tested and the reference product and the proportion in the sample whose test results are positive for the reference product shall be more than 85%.

3)Total coincidence rate: the sample whose test results are the same for the product tested and the reference product and its proportion in the total number of samples shall be more than 90%.

Method		2019-nCoV nucleic acid test kit (RT-PCR)		Total Results
The Novel Coronavirus (SARS-CoV-2) Antigen Rapid Test Cassette(Swab)	Result	positive	negative	
	positive	A	B	A+B
	negative	C	D	C+D
Total Results		A+C	B+D	A+B+C+D

Clinical sensitivity = $A/(A+C)*100\%$

Clinical specificity = $D/(B+D)*100\%$

Accuracy:  $(A+D)/(A+B+C+D)*100\%$

If the coincidence rate of positive/negative can meet clinical requirements, two methods or Products are considered as equivalent; If the coincidence rate of positive/negative is greatly different, the clinical scheme should be re-designed.

4)Kappa consistency analysis shall be adopted for statistical analysis of reference reagents.

The results of the product tested are statistical materials and can be per the table below:

Method	2019-nCoV nucleic acid	Total Results
--------	------------------------	---------------

File No. MF-K511416D-0010

Version: 1.5

Effective date:2020-11-25

		test kit (RT-PCR)		
The Novel Coronavirus (SARS-Cov-2) Antigen Rapid Test Cassette(Swab)	Result	positive	negative	
	positive	A	B	A+B
	negative	C	D	C+D
Total Results		A+C	B+D	A+B+C+D

$$P_0 = (A+D)/(A+B+C+D)*100\%$$

$$P_e = ((A+B)(A+C) + (A+B)(B+D)) / (A+B+C+D)^2$$

$$\text{Kappa} = (P_0 - P_e) / (1 - P_e)$$

If conducting Kappa consistency analysis for the base data above, high consistency can be judged if the Kappa coefficient is  $>0.8$ , and both systems are considered as equivalent. Consistency is considered if  $0.4 < \text{Kappa coefficient} < 0.8$ , and the coincidence rate of positive/negative shall be compared, with statistical analysis being made. Two such systems are considered as inconsistent and in-equivalent if the Kappa coefficient is  $<0.4$ .

### VIII Provisions for amendments to clinical validation

In general, the clinical validation should not be changed. Any modification to the project during the test should be explained, and the time, reason, process of change, and whether there is a record of the change are explained in detail and its impact on the evaluation of the entire research result is explained.

### IX. Results and Analysis of Clinical Tests

In total, 659 test samples are included for the unit and all test samples included are tested. There are 201 positive test samples, among them, 55 test samples have Ct values  $>30$ .

Days from diagnosis	Positive	Negative	Total Number Tested	Detectable rate
0-3	160	0	160	100%
4-7	19	2	21	90.48%
> 7	22	6	28	84.62%
Total	201	8	209	/

Statistics on test results and those of the product tested are as follows:

Method	2019-nCoV Nucleic Acid Test Kit (RT-PCR)			Total Results
	Results	Positive	Negative	
The Novel Coronavirus (SARS-Cov-2) Antigen Rapid Test Cassette (Swab)	Positive	201	0	201
	Negative	8	450	458
	Total Results	209	450	659

$$\text{Clinical sensitivity} = 201/209 = 96.17\% \text{ (95\%CI* 92.51\% to 98.17\%)}$$

$$\text{Clinical specificity} = 450/450 > 99.9\% \text{ (95\%CI* 98.98\% to 100\%)}$$

File No. MF-K511416D-0010

Version: 1.5

Effective date: 2020-11-25

Accuracy:  $(201+450)/(201+0+8+450) * 100\% = 98.79\%$  (95%CI\* 97.58% to 99.43%)

$P_e = (209*450+458*201)/(659*659) = 0.57$

Kappa:  $(P_0 - P_e)/(1 - P_e) = 0.97$

\*: 95% confidence interval

According to the above table, 450 are proven negative of 450 negative specimens, 201 are proven positive of 209 positive specimens. The sensitivity and accuracy are more than 90%, indicating favorable consistency with the reference product. The Kappa=0.97>0.8, indicating favorable and high consistency of two methods and equivalence of two such systems.

#### X Analysis on consistency in Test Results

Consistency result number	RdRP gene Ct/Cq value			N gene Ct/Cq value		
	Maximum	Minimum	Median	Maximum	Minimum	Median
201	34.86	14.01	22.9	35.46	14.01	23.09

According to the test result, there are 201 samples have the consistency results for rapid test and RT-PCR. For qualitative rapid test, the result will show positive, negative and invalid, for RT-PCR detection, the Ct/Cq value will indicate the result, Ct/Cq value > 40 means the detection result is negative, in our validation test, the amount of both RdRP and N gene Ct/Cq value are all below 40 is 201. The median of RdRP gene Ct/Cq value is 22.9, while the RdRP gene Ct/Cq value is 23.09.

#### XI Analysis on Inconsistency in Test Results

NO.	Age	Gender	Rapid Test	Ct/Cq value (RT-PCR)		Clinical diagnostic
				RdRP	N gene	
9	23	F	Negative	> 40	34.38	Infection 25 days
49	30	M	Negative	> 40	31.74	Infection 16 days
72	39	M	Negative	34.76	> 40	Infection 28 days
80	28	M	Negative	30.62	31.23	Infection 31 days
105	58	F	Negative	> 40	35.27	Infection 19 days
147	52	M	Negative	> 40	31.37	Infection 22 days
154	16	F	Negative	> 40	35.01	Infection 7* days
209	67	F	Negative	32.60	33.25	Infection 5* days

\*There are two mistakes when we checked the sample information again, the true days from diagnosis for sample 154 and sample 209 are 7 and 5 days.

### XII Discussion and Conclusions

#### 1. discussion

A Results of comparative analysis of the product tested and the reference product:

Test results of Swab specimen tested and the reference result: both the coincidence rate of negative/positive and the total coincidence rate are larger than 85%, indicating favorable consistency with the reference product. In the analysis results of Kappa inspection, Kappa was proven >0.8, indicating favorable and high consistency of both methods. Both systems were proven equivalent.

#### 2. Test conclusions

By analyzing the test results of the product tested and the reference product, the consistency percentage of negative/positive and the total consistency percentage are proven high. Moreover, according to the results of statistical analysis, there is no remarkable difference in test results of

File No. MF-K511416D-0010

Version: 1.5

Effective date: 2020-11-25

both, indicating favorable consistency in diagnosis and equivalence of two such systems and can be used for auxiliary diagnosis of those suffering from pneumonia triggered by COVID-19.

### **XIII. Quality control methods**

#### On-site quality control

1) During the course of this study, clinical implementors appointed clinical inspectors to conduct regular on-site supervision visits to the research hospital. Through monitoring visits, it was found that all the contents of the research plan were strictly observed, and the correctness of the research data was also guaranteed. Participating researchers have undergone unified training, unified recording methods and judgment standards. The entire clinical trial process is conducted under strict operation, and the test content is complete and authentic. All observations and findings in the clinical trials have been verified and the data are reliable. The conclusions in the clinical trials are derived from the original data.

2) Quality control of clinical experiment process

During the evaluation, quality control was performed daily to ensure that the product was under control. Strict quality control is performed for each trial to ensure the quality of clinical trials.

### **XIV. Prediction of adverse events**

Because the Novel Coronavirus (SARS-Cov-2) Antigen Rapid Test Cassette (Swab) is an in vitro diagnostic reagent product, no direct contact with patients is required in clinical trials, no test report is provided to patients, and the test results are only used for comparative studies. It involves personal privacy, does not serve as a basis for auxiliary diagnosis, does not bring any risk to the subject, and does not cause adverse events.

#### **References:**

- 1.The "Technical Review Points for the Registration of New Coronavirus Antigen / Antibody Detection Reagents in 2019 (Trial)" issued by the State Drug Administration Medical Device Technical Evaluation Center on February 25, 2020;
2. "Pneumonitis Diagnosis and Treatment Program for New Coronavirus Infection (Trial Version 7)" issued by the National Health Committee on February 19, 2020.



<b>Streptococcus pyogenes</b>	Turing strain T1 NCB#1184; SF130]	1 x 10 <sup>3</sup> PFU/ml
Methicillin 22	1 x 10 <sup>3</sup> PFU/ml	
<b>Mycoplasma pneumoniae</b>	1 x 10 <sup>3</sup> PFU/ml	
MRSA (ATCC 35613)	1 x 10 <sup>3</sup> PFU/ml	
229E	1.5 x 10 <sup>3</sup> TCID <sub>50</sub> /ml	
OC43	1.5 x 10 <sup>3</sup> TCID <sub>50</sub> /ml	
NL63	1.5 x 10 <sup>3</sup> TCID <sub>50</sub> /ml	
HKU1	1.5 x 10 <sup>3</sup> TCID <sub>50</sub> /ml	
Human coronavirus (hCoV) 229E/B1	1.5 x 10 <sup>3</sup> TCID <sub>50</sub> /ml	
Human Measlesvirus (hMPV) 16 Type A1	1.5 x 10 <sup>3</sup> TCID <sub>50</sub> /ml	
<b>Parainfluenza virus</b>	Type 1 ..... Type 2 ..... Type 3 ..... Type 4A ..... Type 4B .....	1.5 x 10 <sup>3</sup> TCID <sub>50</sub> /ml 1.5 x 10 <sup>3</sup> TCID <sub>50</sub> /ml

**Interfering Substances Reaction**

When tested using the Novel Coronavirus (2019-nCoV) Antigen Rapid Test Cassette (swab), there was no interference between the device reagents and the potential interference substances listed in below table that would create false positive or negative results for SARS-CoV-2 antigen.

Substance	Concentration	Concentration
Vitamin	0.05 μM	0.05 μM
Biotin	0.5 μM	2.5 μM
Extrin	0.05 μM	0.5 μM
β-Hydroxybutyrate	100 μM	100 μM
β-Alanide (Dihydrofuran)	100 μM	100 μM
β-Alanide (Glycine)	100 μM	100 μM
β-Alanide (Glycine)	100 μM	100 μM
Saline Nasal Spray	50 μM	50 μM
Homoeopathic	50 μM	50 μM
Sodium Cromoglicate	0.001M	0.001M
Clohexidine Hydrochloride	0.001M	0.001M
Zamanner	5 mg/ml	5 mg/ml
Cetadine	0.001M	0.001M
Artemether-Lumefantrine	50 μM	100 μM
Doxycycline hydroclate	50 μM	100 μM
Guanidine	1 mg/ml	0.64 mM/L
Aminodrine	1 mg/ml	0.29 mM
Fluconazole	1 mg/ml	0.29 mM
Clotrimazole	1 mg/ml	0.29 mM
Carbameze	1 mg/ml	0.173 mM
Acetaminophen	150 μM	NoA

Symbol	Meaning	Symbol	Meaning
[IVD]	In vitro diagnostic medical device	✓	Storage temperature limit
[EC REP]	Authorized representative in the European Community		
[Date]	Date of Manufacture	☒	Use by date
[Do not reuse]	Do not reuse	☒	Cancel instruction for reuse
[LOT]	Batch code	CE	Meet the requirements of EC Directive 93/42/EC



Number: 1101381601  
Version: 0031  
Effective Date: 2020-10-07

File No. MF-K511416D-0010  
Version: 1.5  
Effective date: 2020-11-25

## Annex II: Information of sample

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)
				RdRP	N gene
1	49	F	1	Positive	16.29
2	32	F	2	Positive	14.97
3	31	F	1	Positive	27.24
4	32	F	2	Positive	17.74
5	21	F	3	Positive	21.38
6	51	M	0	Positive	15.43
7	22	F	25	Positive	26.73
8	46	F	1	Positive	21.21
9	23	F	25	Negative	>40
10	14	M	3	positive	34.38
11	42	M	3	Positive	22.65
12	51	M	3	Positive	23.85
13	80	M	2	Positive	24.33
14	39	F	3	Positive	14.65
15	67	M	9	Positive	31.24
16	44	M	1	positive	30.93
17	26	F	20	Positive	22.71
18	33	F	1	positive	31.55
19	38	F	12	Positive	14.73
20	36	F	1	Positive	14.58
21	3	F	5	Positive	34.65

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)
				RdRP	N gene
1	49	F	1	Positive	15.96
2	32	F	2	Positive	23
3	31	F	1	Positive	27.78
4	32	F	2	Positive	24
5	21	F	3	Positive	25
6	51	M	0	Positive	26
7	22	F	25	Positive	27
8	46	F	1	Positive	28
9	23	F	25	Negative	20
10	14	M	3	positive	29
11	42	M	3	Positive	26
12	51	M	3	Positive	27
13	80	M	2	Positive	28
14	39	F	3	Positive	29
15	67	M	9	Positive	30
16	44	M	1	positive	31
17	26	F	20	Positive	31
18	33	F	1	positive	32
19	38	F	12	Positive	31
20	36	F	1	Positive	34
21	3	F	5	Positive	34

**REALY**

杭州睿丽科技有限公司

Hangzhou Realy Tech Co., Ltd.

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)		RdRP	N gene	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)
					RdRP	N gene					
43	75	F	1	Positive	28.64	28.35	66	58	F	0	Positive
44	60	F	2	Positive	21.03	21.45	67	20	F	0	Positive
45	25	M	3	Positive	32.88	32.55	68	42	M	2	Positive
46	75	F	1	Positive	16.82	17.16	69	27	F	1	Positive
47	43	F	5	Positive	18.66	19.03	70	49	M	2	Positive
48	30	F	1	Positive	15.27	15.12	71	49	M	1	positive
49	30	M	16	Negative	>40	31.74	72	39	M	28	Negative
50	26	F	3	Positive	32.61	32.94	73	17	F	7	Positive
51	32	F	1	Positive	24.58	24.09	74	60	M	4	Positive
52	73	M	1	Positive	18.93	19.31	75	44	M	0	Positive
53	58	F	2	Positive	22.91	23.14	76	49	F	2	Positive
54	66	F	1	Positive	33.03	33.03	77	11	M	0	Positive
55	29	F	3	Positive	17.96	17.78	78	32	M	3	positive
56	56	M	3	Positive	31.67	32.3	79	51	F	1	Positive
57	24	M	29	Positive	21.89	21.67	80	28	M	31	Negative
58	36	M	2	Positive	20.6	20.19	81	31	F	3	Positive
59	70	F	2	Positive	16.65	16.82	82	50	M	0	Positive
60	45	M	7	Positive	24.93	24.68	83	47	M	28	Positive
61	38	F	3	Positive	34.76	35.46	84	44	F	2	Positive
62	42	M	1	Positive	32.69	32.04	85	10	F	26	Positive
63	55	M	0	Positive	20.86	21.28	86	24	M	15	Positive
64	33	M	0	Positive	21.06	20.64	87	22	F	5	Positive
65	39	M	1	Positive	24.32	24.08	88	47	F	1	Positive

**REALY**

杭州睿丽科技有限公司  
Hangzhou Realy Tech Co., Ltd.

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)		RdRP	N gene
					RdRP	N gene		
89	26	M	1	Positive	17.3	16.95	112	15
90	43	F	2	Positive	33.26	32.59	113	53
91	55	M	7	Positive	14.83	15.13	114	10
92	50	F	0	Positive	16.85	17.02	115	25
93	26	M	3	Positive	25.72	25.72	116	39
94	51	F	2	Positive	19.19	19.57	117	56
95	20	F	0	Positive	23.4	23.63	118	49
96	44	M	7	Positive	31.19	31.81	119	20
97	38	F	1	Positive	34.86	34.51	120	25
98	38	F	20	Positive	17.09	17.09	121	37
99	39	M	2	Positive	29.32	28.73	122	52
100	30	F	2	Positive	27.24	27.24	123	60
101	57	F	0	Positive	28.67	29.24	124	25
102	45	M	3	Positive	22.36	22.81	125	19
103	41	F	3	Positive	17.74	17.92	126	32
104	26	F	2	Positive	27.7	27.15	127	28
105	58	F	19	Negative	>40	35.25	128	52
106	39	F	3	Positive	15.08	15.38	129	40
107	60	F	0	Positive	24.62	24.37	130	28
108	11	M	5	Positive	20.59	20.38	131	31
109	12	F	3	Positive	25.89	26.41	132	48
110	17	M	3	Positive	19.5	19.5	133	33
111	59	F	7	Positive	23.46	22.99	134	44

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)		RdRP	N gene
					RdRP	N gene		
112	15	M	3	Positive	29.78	29.78		
113	53	F	1	Positive	14.37	14.08		
114	10	M	1	Positive	19.57	19.57		
115	25	F	2	Positive	23.88	23.64		
116	39	M	3	Positive	28.43	28.71		
117	56	M	5	Positive	30.88	30.88		
118	49	M	5	Positive	19.61	19.81		
119	20	M	3	Positive	20.95	20.74		
120	25	M	0	Positive	23.78	23.3		
121	37	F	1	Positive	21.99	21.77		
122	52	F	3	Positive	27.08	26.54		
123	60	F	23	Positive	24.66	24.91		
124	25	M	2	Positive	23.93	23.45		
125	19	F	2	Positive	33.72	33.38		
126	32	M	0	Positive	25.86	25.86		
127	28	F	27	Positive	32.49	33.14		
128	52	F	6	Positive	20.82	20.4		
129	40	F	2	Positive	31.45	30.82		
130	28	F	2	Positive	31.43	31.74		
131	31	F	1	Positive	22.3	21.85		
132	48	M	2	Positive	27.47	27.2		
133	33	F	3	positive	23.31	23.08		
134	44	M	12	positive	15.2	15.35		

**REAl**

杭州睿丽科技有限公司

Hangzhou Realy Tech Co., Ltd.

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)		Rapid Test	Ct/Cq value (RT-PCR)
					RdRP	N gene		
135	34	F	29	Positive	19.96	20.16	158	24
136	18	M	1	Positive	32.05	31.73	159	36
137	59	M	3	Positive	18.8	18.61	160	19
138	18	M	14	Positive	17.2	17.54	161	33
139	38	M	0	Positive	34.82	34.47	162	53
140	20	F	1	Positive	22.51	22.96	163	43
141	54	F	1	Positive	28.96	28.38	164	12
142	43	F	0	Positive	14.54	14.25	165	14
143	23	M	3	Positive	24.58	24.09	166	56
144	27	F	0	Positive	23.3	23.53	167	52
145	39	M	1	Positive	21.94	21.94	168	32
146	60	M	7	Positive	25.1	25.35	169	50
147	52	M	22	Negative	>40	31.37	170	18
148	49	M	2	Positive	19.68	19.29	171	35
149	42	M	1	positive	14.85	15	172	12
150	32	F	1	Positive	24.91	24.91	173	14
151	59	M	3	Positive	25.69	25.69	174	13
152	33	F	2	Positive	25.71	25.2	175	42
153	15	F	2	Positive	27.12	26.85	176	13
154	16	F	7	Negative	>40	35.01	177	14
155	24	M	2	Positive	23.46	23.46	178	60
156	52	F	1	Positive	14.01	14.01	179	13
157	60	M	3	Positive	27.28	27.28	180	51

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)		Rapid Test	Ct/Cq value (RT-PCR)
					RdRP	N gene		
158	24	M	3	Positive	21.21	21		
159	36	F	1	Positive	16.23	15.91		
160	19	F	3	Positive	21.11	21.32		
161	33	F	0	Positive	17.7	17.52		
162	53	M	0	Positive	18.99	19.37		
163	43	M	3	Positive	14.46	14.75		
164	12	F	2	Positive	34.25	33.91		
165	14	M	1	Positive	14.68	14.68		
166	56	M	3	Positive	16.37	16.7		
167	52	F	0	Positive	26.62	26.89		
168	32	F	3	Positive	15.2	15.5		
169	50	M	0	Positive	32.2	32.52		
170	18	F	2	Positive	18.52	18.15		
171	35	F	2	Positive	28.17	27.89		
172	12	M	0	Positive	28.99	28.99		
173	14	F	27	Positive	32.57	32.9		
174	13	F	3	Positive	27.88	28.16		
175	42	M	3	Positive	16.8	16.46		
176	13	F	0	Positive	20.05	19.65		
177	14	F	3	Positive	19.49	19.49		
178	60	M	0	Positive	28.7	29.27		
179	13	F	0	Positive	27.63	27.91		
180	51	M	1	Positive	21.79	22.23		

**REAlY**

杭州睿丽科技有限公司

**Hangzhou Realy Tech Co., Ltd.**

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)		RdRP	N gene	RdRP	N gene
					RdRP	N gene				
181	56	M	3	Positive	17.33	17.5	204	M	6	Positive
182	14	F	3	Positive	23.94	23.46	205	M	2	Positive
183	45	F	0	Positive	32.37	32.05	206	M	14	Positive
184	24	F	1	Positive	20.03	20.03	207	F	1	Positive
185	15	M	2	Positive	31.31	31	208	M	1	Positive
186	51	F	7	Positive	15.26	15.11	209	F	5	Negative
187	31	M	2	Positive	27.51	27.51	210	F	/	Negative
188	49	M	7	Positive	16.07	16.23	211	M	/	Negative
189	28	M	0	Positive	27.15	26.61	212	F	/	Negative
190	80	M	0	Positive	31.13	31.13	213	F	/	Negative
191	47	M	20	Positive	17.21	17.21	214	M	/	Negative
192	22	F	3	Positive	31.33	31.33	215	F	/	Negative
193	49	F	24	Positive	25.13	25.38	216	M	/	Negative
194	23	M	3	Positive	18.2	17.84	217	M	/	Negative
195	30	F	1	Positive	21.4	21.4	218	F	/	Negative
196	55	F	7	Positive	27.42	27.69	219	F	/	Negative
197	75	F	0	Positive	22.87	22.87	220	F	/	Negative
198	49	M	0	Positive	14.63	14.78	221	M	/	Negative
199	81	M	24	Positive	20.42	20.22	222	F	/	Negative
200	51	F	0	Positive	22.09	22.53	223	M	/	Negative
201	12	F	3	Positive	16.93	17.1	224	M	/	Negative
202	47	M	2	Positive	27.4	27.13	225	M	/	Negative
203	78	F	1	Positive	18.69	18.69	226	F	/	Negative

**REAl**

杭州睿丽科技有限公司

Hangzhou Realy Tech Co., Ltd.

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)		RdRP	N gene	Ct/Cq value (RT-PCR)
					RdRP	N gene			
227	21	F	/	Negative	>40	>40	250	68	M
228	46	F	/	Negative	>40	>40	251	29	M
229	71	M	/	Negative	>40	>40	252	54	M
230	60	F	/	Negative	>40	>40	253	49	M
231	31	F	/	Negative	>40	>40	254	20	M
232	72	M	/	Negative	>40	>40	255	26	M
233	62	M	/	Negative	>40	>40	256	22	M
234	39	F	/	Negative	>40	>40	257	32	F
235	45	M	/	Negative	>40	>40	258	28	M
236	21	M	/	Negative	>40	>40	259	44	M
237	33	M	/	Negative	>40	>40	260	57	F
238	83	M	/	Negative	>40	>40	261	64	F
239	15	M	/	Negative	>40	>40	262	39	F
240	59	M	/	Negative	>40	>40	263	38	F
241	54	M	/	Negative	>40	>40	264	73	M
242	84	F	/	Negative	>40	>40	265	45	M
243	84	F	/	Negative	>40	>40	266	61	M
244	42	F	/	Negative	>40	>40	267	13	F
245	63	F	/	Negative	>40	>40	268	64	F
246	29	M	/	Negative	>40	>40	269	26	F
247	50	M	/	Negative	>40	>40	270	28	M
248	74	F	/	Negative	>40	>40	271	58	M
249	43	M	/	Negative	>40	>40	272	35	F

**REALY**

杭州睿丽科技有限公司

Hangzhou Realy Tech Co., Ltd.

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)			Rapid Test	Ct/Cq value (RT-PCR)
					RdRP	N gene	RdRP		
273	51	M	/	Negative	>40	>40	296	58	F
274	60	M	/	Negative	>40	>40	297	68	M
275	17	M	/	Negative	>40	>40	298	77	M
276	18	F	/	Negative	>40	>40	299	47	F
277	15	M	/	Negative	>40	>40	300	71	M
278	52	M	/	Negative	>40	>40	301	21	F
279	33	M	/	Negative	>40	>40	302	52	M
280	41	F	/	Negative	>40	>40	303	70	M
281	11	M	/	Negative	>40	>40	304	63	M
282	19	F	/	Negative	>40	>40	305	59	M
283	10	F	/	Negative	>40	>40	306	26	M
284	62	F	/	Negative	>40	>40	307	36	F
285	68	F	/	Negative	>40	>40	308	47	F
286	38	M	/	Negative	>40	>40	309	45	M
287	59	M	/	Negative	>40	>40	310	29	F
288	76	F	/	Negative	>40	>40	311	30	M
289	24	M	/	Negative	>40	>40	312	25	F
290	68	M	/	Negative	>40	>40	313	73	M
291	82	F	/	Negative	>40	>40	314	76	M
292	64	F	/	Negative	>40	>40	315	25	M
293	59	M	/	Negative	>40	>40	316	49	F
294	59	M	/	Negative	>40	>40	317	62	M
295	83	M	/	Negative	>40	>40	318	38	M

REAlY

杭州睿丽科技有限公司  
Hangzhou Realy Tech Co., Ltd.

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)		RdRP	N gene	Ct/Cq value (RT-PCR)
					RdRP	N gene			
319	33	M	/	Negative	>40	>40	342	/	Negative
320	39	M	/	Negative	>40	>40	343	/	Negative
321	69	M	/	Negative	>40	>40	344	/	Negative
322	79	F	/	Negative	>40	>40	345	/	Negative
323	32	M	/	Negative	>40	>40	346	/	Negative
324	35	M	/	Negative	>40	>40	347	/	Negative
325	39	M	/	Negative	>40	>40	348	/	Negative
326	61	F	/	Negative	>40	>40	349	/	Negative
327	10	F	/	Negative	>40	>40	350	/	Negative
328	37	M	/	Negative	>40	>40	351	/	Negative
329	52	F	/	Negative	>40	>40	352	/	Negative
330	41	M	/	Negative	>40	>40	353	/	Negative
331	74	M	/	Negative	>40	>40	354	/	Negative
332	51	F	/	Negative	>40	>40	355	/	Negative
333	56	M	/	Negative	>40	>40	356	/	Negative
334	62	F	/	Negative	>40	>40	357	/	Negative
335	60	F	/	Negative	>40	>40	358	/	Negative
336	54	F	/	Negative	>40	>40	359	/	Negative
337	81	F	/	Negative	>40	>40	360	/	Negative
338	79	F	/	Negative	>40	>40	361	/	Negative
339	73	F	/	Negative	>40	>40	362	/	Negative
340	35	F	/	Negative	>40	>40	363	/	Negative
341	76	F	/	Negative	>40	>40	364	/	Negative

**REAl**

杭州睿丽科技有限公司

**Hangzhou Realy Tech Co., Ltd.**

NO.	Age	Gender	Days from diagnose	Rapid Test		Ct/Cq value (RT-PCR)		RdRP	N gene	Ct/Cq value (RT-PCR)
				RdRP	N gene	RdRP	N gene			
365	37	F	/	Negative	>40	>40	/	Negative	>40	>40
366	63	M	/	Negative	>40	>40	/	Negative	>40	>40
367	39	F	/	Negative	>40	>40	M	/	Negative	>40
368	38	M	/	Negative	>40	>40	/	Negative	>40	>40
369	37	M	/	Negative	>40	>40	/	Negative	>40	>40
370	37	F	/	Negative	>40	>40	/	Negative	>40	>40
371	56	F	/	Negative	>40	>40	M	/	Negative	>40
372	56	F	/	Negative	>40	>40	M	/	Negative	>40
373	59	M	/	Negative	>40	>40	/	Negative	>40	>40
374	13	M	/	Negative	>40	>40	M	/	Negative	>40
375	80	F	/	Negative	>40	>40	/	Negative	>40	>40
376	59	M	/	Negative	>40	>40	M	/	Negative	>40
377	61	F	/	Negative	>40	>40	M	/	Negative	>40
378	70	M	/	Negative	>40	>40	M	/	Negative	>40
379	20	M	/	Negative	>40	>40	M	/	Negative	>40
380	75	F	/	Negative	>40	>40	M	/	Negative	>40
381	49	M	/	Negative	>40	>40	F	/	Negative	>40
382	47	M	/	Negative	>40	>40	F	/	Negative	>40
383	65	F	/	Negative	>40	>40	M	/	Negative	>40
384	78	M	/	Negative	>40	>40	F	/	Negative	>40
385	84	M	/	Negative	>40	>40	M	/	Negative	>40
386	72	F	/	Negative	>40	>40	M	/	Negative	>40
387	20	F	/	Negative	>40	>40	M	/	Negative	>40



杭州睿丽科技有限公司

Hangzhou Realy Tech Co., Ltd.

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)	RdRP	N gene
411	29	F	/	Negative	>40	>40	>40
412	63	F	/	Negative	>40	>40	>40
413	56	M	/	Negative	>40	>40	>40
414	28	M	/	Negative	>40	>40	>40
415	50	F	/	Negative	>40	>40	>40
416	21	F	/	Negative	>40	>40	>40
417	24	M	/	Negative	>40	>40	>40
418	51	F	/	Negative	>40	>40	>40
419	63	M	/	Negative	>40	>40	>40
420	22	M	/	Negative	>40	>40	>40
421	55	F	/	Negative	>40	>40	>40
422	11	F	/	Negative	>40	>40	>40
423	37	F	/	Negative	>40	>40	>40
424	60	F	/	Negative	>40	>40	>40
425	78	M	/	Negative	>40	>40	>40
426	48	M	/	Negative	>40	>40	>40
427	39	M	/	Negative	>40	>40	>40
428	31	F	/	Negative	>40	>40	>40
429	24	M	/	Negative	>40	>40	>40
430	51	F	/	Negative	>40	>40	>40
431	43	M	/	Negative	>40	>40	>40
432	49	F	/	Negative	>40	>40	>40
433	18	F	/	Negative	>40	>40	>40

NO.	Age	Gender	Days from diagnose	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)	RdRP	N gene
434	32	M	/			Negative	>40	>40	
435	77	M	/			Negative	>40	>40	
436	47	M	/			Negative	>40	>40	
437	82	F	/			Negative	>40	>40	
438	38	F	/			Negative	>40	>40	
439	51	M	/			Negative	>40	>40	
440	40	F	/			Negative	>40	>40	
441	21	F	/			Negative	>40	>40	
442	60	M	/			Negative	>40	>40	
443	80	F	/			Negative	>40	>40	
444	12	M	/			Negative	>40	>40	
445	68	F	/			Negative	>40	>40	
446	11	M	/			Negative	>40	>40	
447	55	M	/			Negative	>40	>40	
448	83	M	/			Negative	>40	>40	
449	83	M	/			Negative	>40	>40	
450	84	F	/			Negative	>40	>40	
451	29	F	/			Negative	>40	>40	
452	53	F	/			Negative	>40	>40	
453	42	M	/			Negative	>40	>40	
454	48	M	/			Negative	>40	>40	
455	34	F	/			Negative	>40	>40	
456	40	M	/			Negative	>40	>40	

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)	RdRP	N gene
457	77	F	/	Negative	>40	>40	>40
458	39	F	/	Negative	>40	>40	>40
459	81	M	/	Negative	>40	>40	>40
460	63	M	/	Negative	>40	>40	>40
461	15	M	/	Negative	>40	>40	>40
462	81	F	/	Negative	>40	>40	>40
463	79	M	/	Negative	>40	>40	>40
464	58	M	/	Negative	>40	>40	>40
465	23	M	/	Negative	>40	>40	>40
466	15	M	/	Negative	>40	>40	>40
467	82	M	/	Negative	>40	>40	>40
468	48	M	/	Negative	>40	>40	>40
469	73	F	/	Negative	>40	>40	>40
470	71	M	/	Negative	>40	>40	>40
471	69	F	/	Negative	>40	>40	>40
472	22	M	/	Negative	>40	>40	>40
473	52	M	/	Negative	>40	>40	>40
474	26	M	/	Negative	>40	>40	>40
475	82	M	/	Negative	>40	>40	>40
476	36	M	/	Negative	>40	>40	>40
477	46	M	/	Negative	>40	>40	>40
478	47	F	/	Negative	>40	>40	>40
479	24	F	/	Negative	>40	>40	>40

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)	RdRP	N gene
480	33	M	/	Negative	>40	>40	>40
481	17	M	/	Negative	>40	>40	>40
482	34	F	/	Negative	>40	>40	>40
483	76	F	/	Negative	>40	>40	>40
484	53	M	/	Negative	>40	>40	>40
485	53	M	/	Negative	>40	>40	>40
486	76	F	/	Negative	>40	>40	>40
487	66	F	/	Negative	>40	>40	>40
488	57	F	/	Negative	>40	>40	>40
489	21	F	/	Negative	>40	>40	>40
490	35	M	/	Negative	>40	>40	>40
491	21	F	/	Negative	>40	>40	>40
492	21	M	/	Negative	>40	>40	>40
493	28	F	/	Negative	>40	>40	>40
494	58	M	/	Negative	>40	>40	>40
495	37	M	/	Negative	>40	>40	>40
496	22	M	/	Negative	>40	>40	>40
497	65	M	/	Negative	>40	>40	>40
498	29	M	/	Negative	>40	>40	>40
499	48	M	/	Negative	>40	>40	>40
500	11	M	/	Negative	>40	>40	>40
501	29	F	/	Negative	>40	>40	>40
502	11	F	/	Negative	>40	>40	>40

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)
				RdRP	N gene
503	79	M	/	Negative	>40
504	46	F	/	Negative	>40
505	14	M	/	Negative	>40
506	17	M	/	Negative	>40
507	72	F	/	Negative	>40
508	83	F	/	Negative	>40
509	30	F	/	Negative	>40
510	71	M	/	Negative	>40
511	79	M	/	Negative	>40
512	84	F	/	Negative	>40
513	62	M	/	Negative	>40
514	50	F	/	Negative	>40
515	21	F	/	Negative	>40
516	81	F	/	Negative	>40
517	76	F	/	Negative	>40
518	41	F	/	Negative	>40
519	73	M	/	Negative	>40
520	83	F	/	Negative	>40
521	71	M	/	Negative	>40
522	10	M	/	Negative	>40
523	63	M	/	Negative	>40
524	72	M	/	Negative	>40
525	59	M	/	Negative	>40

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)
				RdRP	N gene
526	35	M	/	Negative	>40
527	58	M	/	Negative	>40
528	46	F	/	Negative	>40
529	79	M	/	Negative	>40
530	76	M	/	Negative	>40
531	77	F	/	Negative	>40
532	45	F	/	Negative	>40
533	73	M	/	Negative	>40
534	38	F	/	Negative	>40
535	41	F	/	Negative	>40
536	32	F	/	Negative	>40
537	50	M	/	Negative	>40
538	31	M	/	Negative	>40
539	74	F	/	Negative	>40
540	16	F	/	Negative	>40
541	69	M	/	Negative	>40
542	72	M	/	Negative	>40
543	40	F	/	Negative	>40
544	78	F	/	Negative	>40
545	53	M	/	Negative	>40
546	44	F	/	Negative	>40
547	28	M	/	Negative	>40
548	14	M	/	Negative	>40

NO.	Age	Gender	Days from diagnose	Rapid Test		Ct/Cq value (RT-PCR)		Rapid Test	Ct/Cq value (RT-PCR)
				RdRP	N gene	RdRP	N gene		
549	80	F	/	Negative	>40	>40	>40	/	Negative
550	40	F	/	Negative	>40	>40	>40	/	Negative
551	26	M	/	Negative	>40	>40	>40	/	Negative
552	11	F	/	Negative	>40	>40	>40	/	Negative
553	53	F	/	Negative	>40	>40	>40	/	Negative
554	19	M	/	Negative	>40	>40	>40	/	Negative
555	21	M	/	Negative	>40	>40	>40	/	Negative
556	60	M	/	Negative	>40	>40	>40	/	Negative
557	12	F	/	Negative	>40	>40	>40	/	Negative
558	58	M	/	Negative	>40	>40	>40	/	Negative
559	62	M	/	Negative	>40	>40	>40	/	Negative
560	45	M	/	Negative	>40	>40	>40	/	Negative
561	34	F	/	Negative	>40	>40	>40	/	Negative
562	35	F	/	Negative	>40	>40	>40	/	Negative
563	82	M	/	Negative	>40	>40	>40	/	Negative
564	59	F	/	Negative	>40	>40	>40	/	Negative
565	59	F	/	Negative	>40	>40	>40	/	Negative
566	38	F	/	Negative	>40	>40	>40	/	Negative
567	82	M	/	Negative	>40	>40	>40	/	Negative
568	22	F	/	Negative	>40	>40	>40	/	Negative
569	50	M	/	Negative	>40	>40	>40	/	Negative
570	25	M	/	Negative	>40	>40	>40	/	Negative
571	52	M	/	Negative	>40	>40	>40	/	Negative

NO.	Age	Gender	Days from diagnose	Rapid Test		Ct/Cq value (RT-PCR)		Rapid Test	Ct/Cq value (RT-PCR)
				RdRP	N gene	RdRP	N gene		
572	13	M	/	/	/	/	/	>40	>40
573	33	M	/	/	/	/	/	>40	>40
574	60	F	/	/	/	/	/	>40	>40
575	43	F	/	/	/	/	/	>40	>40
576	46	M	/	/	/	/	/	>40	>40
577	76	F	/	/	/	/	/	>40	>40
578	34	M	/	/	/	/	/	>40	>40
579	52	F	/	/	/	/	/	>40	>40
580	50	F	/	/	/	/	/	>40	>40
581	64	F	/	/	/	/	/	>40	>40
582	52	M	/	/	/	/	/	>40	>40
583	57	M	/	/	/	/	/	>40	>40
584	50	F	/	/	/	/	/	>40	>40
585	52	M	/	/	/	/	/	>40	>40
586	60	F	/	/	/	/	/	>40	>40
587	16	F	/	/	/	/	/	>40	>40
588	18	F	/	/	/	/	/	>40	>40
589	58	M	/	/	/	/	/	>40	>40
590	26	F	/	/	/	/	/	>40	>40
591	62	F	/	/	/	/	/	>40	>40
592	28	M	/	/	/	/	/	>40	>40
593	50	M	/	/	/	/	/	>40	>40
594	26	M	/	/	/	/	/	>40	>40



杭州睿丽科技有限公司

Hangzhou Realy Tech Co., Ltd.

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)
				RdRP	N gene
595	82	F	/	Negative	>40
596	24	F	/	Negative	>40
597	77	M	/	Negative	>40
598	13	M	/	Negative	>40
599	62	M	/	Negative	>40
600	47	M	/	Negative	>40
601	62	M	/	Negative	>40
602	33	F	/	Negative	>40
603	37	F	/	Negative	>40
604	60	F	/	Negative	>40
605	70	M	/	Negative	>40
606	30	F	/	Negative	>40
607	23	M	/	Negative	>40
608	23	M	/	Negative	>40
609	70	M	/	Negative	>40
610	41	F	/	Negative	>40
611	50	M	/	Negative	>40
612	26	F	/	Negative	>40
613	22	F	/	Negative	>40
614	44	M	/	Negative	>40
615	79	F	/	Negative	>40
616	64	F	/	Negative	>40
617	83	F	/	Negative	>40

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)
				RdRP	N gene
618	76	M	/	Negative	>40
619	25	M	/	Negative	>40
620	41	M	/	Negative	>40
621	30	F	/	Negative	>40
622	30	M	/	Negative	>40
623	37	F	/	Negative	>40
624	46	F	/	Negative	>40
625	48	F	/	Negative	>40
626	20	F	/	Negative	>40
627	77	F	/	Negative	>40
628	55	F	/	Negative	>40
629	55	F	/	Negative	>40
630	80	F	/	Negative	>40
631	45	F	/	Negative	>40
632	17	F	/	Negative	>40
633	47	F	/	Negative	>40
634	48	F	/	Negative	>40
635	30	F	/	Negative	>40
636	55	F	/	Negative	>40
637	16	F	/	Negative	>40
638	43	F	/	Negative	>40
639	35	F	/	Negative	>40
640	67	F	/	Negative	>40

**REAlY**

杭州睿丽科技有限公司

Hangzhou Realy Tech Co., Ltd.

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)
				RdRP	N gene
641	82	F	/	Negative	>40
642	55	F	/	Negative	>40
643	75	F	/	Negative	>40
644	56	F	/	Negative	>40
645	16	F	/	Negative	>40
646	21	F	/	Negative	>40
647	18	F	/	Negative	>40
648	20	F	/	Negative	>40
649	63	F	/	Negative	>40
650	49	F	/	Negative	>40

NO.	Age	Gender	Days from diagnose	Rapid Test	Ct/Cq value (RT-PCR)
				RdRP	N gene
651	63	F	/	Negative	>40
652	18	F	/	Negative	>40
653	27	F	/	Negative	>40
654	29	F	/	Negative	>40
655	47	F	/	Negative	>40
656	26	F	/	Negative	>40
657	65	F	/	Negative	>40
658	28	F	/	Negative	>40
659	67	F	/	Negative	>40



Director: 张晓伟  
Date: 2020.11.25  
Seal of company signature